

## Examination Glove Nitrile Top



EN ISO 374-1:2016 Type B EN ISO 374-5:2016



JKT



VIRUS

<b>Description</b>	Tear-resistant Nitrile-Examination glove Length: 24 cm Wall Thickness: ca. 0,13 mm ~ 5,5 mil AQL: 1,5, non-sterile Weight Size M: 7,0 - 7,2 g/piece
<b>Material</b>	Nitrile
<b>Application areas</b>	Medicine, agriculture, cleaning, light chemical protection
<b>Produktvorteile</b>	Durable and tear-proof ambidextrous food safe latex- and powderfree free from allergenic thiurams, thiazoles and latex proteins
<b>Single packaging</b>	100 pieces per dispenser box
<b>Standards</b>	KAT III, EN ISO 374-1:2016, EN ISO 374-5:2016 (EU-Regulation 2016/425) EN 455/1-4 (EU-Regulation 2017/745 Class 1 medical device)

### Logistical Data:

Item no.	Size	Barcode	Minimum order quantity	Packaging Unit	Pal
15390	S / 6,5-7	4018653153901	1 Pack	1/10	360
15391	M / 7,5-8	4018653153918	1 Pack	1/10	360
15392	L / 8,5-9	4018653153925	1 Pack	1/10	360
15393	XL / 9,5-10	4018653153932	1 Pack	1/10	360
15394	XXL / 10,5-11	4018653153949	1 Pack	1/10	360

**Certification:**

**A) Chemical resistance according to EN ISO 374-1:2016, Type B, EN 374-4:2013**

ID Code	Substance	CAS-Nr.	Class	Performance level*	Degradation
J	n-Heptane	142-82-5	Aliphatic hydrocarbon	3	38,3 %
K	40% Sodium Hydroxide	1310-73-2	Inorganic base	6	-6,8 %
T	37% Formaldehyde	50-00-0	Aldehyde	6	23,7 %

\*legend:

Performance level	1	2	3	4	5	6
Breakthrough time (min)	>10	>30	>60	>120	>240	>480

**B) Microbial resistance according to EN ISO 374-5:2016**

Protection against bacteria and fungal spores: pass

Protection against virus: pass

**C) Medical device according to EN 455/1-4**

SRN:	DE-MF-000007708
Basic UDI-DI:	401865300548